


Malaria Generic Pan-Pf RDT Results Guide



NEGATIVE RESULT




Negative

POSITIVE RESULTS



P. falciparum



Non-falciparum (*P. vivax*, *P. ovale*,
P. malariae or a mixed infection of these)



P. falciparum




P. falciparum monoinfection
or 'Mixed' Infection

NOTE: Test is positive even if the line in the test window is faint.

INVALID RESULTS



No Control Line



No Control Line

Repeat the test using a new RDT if no control line appears.