


Guide to Interpretation of Test Results


Paracheck Pf Device Rapid Test for *P. falciparum* Malaria (Ver.3) (30301025)



NEGATIVE RESULT



POSITIVE RESULTS



Plasmodium falciparum

Note: Test is positive even if the test line is faint.

INVALID RESULTS



No Control Line

Note: Repeat test using a new RDT if no control line appears.