



FIND Evaluation of RapiGEN Inc.

BIOCREDIT COVID-19 Ag

External Report

Version 1.0, 16 October 2020

Copyright and use of the report

Copyright in this report is the property of FIND (or controlled by FIND). You are free to share, copy and redistribute the material in any medium or format provided that:

- (i) attribution: you must give appropriate credit to FIND and indicate if changes were made, you may do so in any reasonable manner, but not in any way that suggests that FIND endorses you or your use;
- (ii) non-commercial: you may not use the report for commercial purposes; and
- (iii) no derivatives: if you remix, transform, or build upon the materials or report, you may not distribute the modified materials or report unless with express authorization from FIND.

Presentation of data on our website does not impact any data ownership rights and FIND is not responsible for any use by any third party of these data. Data sources are provided.

Evaluation Process – private sector engagement

FIND is a non-for-profit foundation, whose mission is to find diagnostic solutions to overcome diseases of poverty in lower- and middle-income countries. It works closely with the private and public sectors and receives funding from donors and some of its industry partners. It has internal fire walls, policies and processes to protect it against any undue influence in its work or the publication of its findings.

More information on our policy and guidelines for working with private sector partners can be found here: <https://www.finddx.org/policies/>

For the COVID-19 response, FIND has commissioned independent evaluations of in vitro diagnostics following an Expression of Interest (EOI) process available on FIND's webpage by which all test submissions were scored according to their regulatory status and time to market; the manufacturing and distribution capacity of the supplier; and the supplier-reported clinical and analytical performance.

Document History

Document Version	Date	Comment
1.0	16 October 2020	Interim version



2 Product Info:

Manufacturer Name	RapiGEN Inc.
Test name	BIOCREDIT COVID-19 Ag
Product Code(s)	G61RHA20
Pack size(s)	20
Contents of kit	Test device sealed in a foil pouch with a desiccant, assay diluent tube, filter cap, sterilized swab, IFU
Equipment and consumables required, but not provided	Equipment: Timer, refrigerator (optional, for storage of specimens prior to testing, if applicable) Consumables: PPE
Product Storage (temperature range)	1-40°C
Shelf-life (months)	24 months
Manufacturing Site (country)	Republic of Korea

3 Study details:

Study design:	Prospective diagnostic evaluation studies across multiple, independent sites to determine the accuracy of COVID-19 antigen RDTs, using consecutive enrolment. Interim analyses are performed at 25% and 50% enrolment, and the evaluation is stopped if tests do not meet 95% specificity. Presence of symptoms, date of symptom onset and hospitalization status is collected for all enrolled participants.
Index assays:	Novel lateral flow format tests that detect recombinant SARS-CoV-2 antigens.
Reference method:	Results of the index test are compared to the routine, diagnostic RT-PCR result, which is used for clinical management
Limit of detection:	Analytical sensitivity, i.e. Limit of detection, was performed at the Liverpool School of Tropical Medicine and/or the Institute of Virology, Charité-Universitätsmedizin Berlin in which standardized serial dilutions of cultured viral isolate were prepared. Dilutions were done in triplicate and the LOD was defined as the last dilution where all repeats were interpreted as positive.
Clinical Performance:	Sensitivity was calculated as the proportion of true positive results detected by RapiGEN BIOCREDIT COVID-19 Ag among all positives by the reference method, and reported as a percentage



	<p>Specificity was calculated as the proportion of true negative specimens, identified as negative by RapiGEN BIOCREDIT COVID-19 Ag among all negatives by the reference method, and reported as a percentage.</p> <p>The 95% confidence intervals were calculated in order to assess the level of uncertainty introduced by sample size, using the Wilson's score method.</p>
Ease of Use	<p>A System usability survey and ease of use questionnaire assessing the quality of the test, test preparation, ease of test execution, procedure time, ease of result interpretation, storage conditions and perceived settings of use was completed by operators and a final score out of 100 was calculated.</p>

4 Evaluation Details

Country of Collaborator	Brazil
Location of clinical site(s) (city, town)	Marica, state of Rio de Janeiro
Health care level of site(s)	Community testing clinic
Study period (date to date)	27 July – 16 September 2020
Study cohort inclusion/exclusion	<p>Adults in community meeting national suspect definition</p> <p>Provided informed consent</p>
Sample type, antigen test	Nasopharyngeal swab
Reference PCR Method	<p>Lab-developed assay based on the US CDC protocol, which targets two regions (N1 and N2) of the nucleocapsid (N) gene of SARS-CoV-2 (https://www.fda.gov/media/134922/download)</p>
Sample type, PCR test	Nasopharyngeal swab

5 Results

5.1 Study Cohort (NOTE: if multiple sites, one column per site/country)

Country	Brazil
Total N (valid PCR results)	476
Age [mean (min-max), N]	44.8 (0-106), 473



Gender [%F, (n/N)]	46.7%, (221/473)
Symptoms present [%Yes, (n/N)]	98.7%, (470/476)
Hospitalized (n, % Yes)	Not available
Days from symptom onset [median (Q1-Q3); N]	5 (4-7); 470
Days < 0-3 (n, %)	95, 20%
Days 4-7 (n, %)	296, 63%
Days 8+ (n, %)	79, 17%
Positivity [%, (n/N)]	25%, (117/476)
PCR Ct [median (Q1-Q3); N]	25.6 (20.4-31.1); 117
Ct > 33 (n, %)	20, 17%
Ct > 30 (n, %)	34, 29%
Ct > 25 (n, %)	62, 53%

5.2 Estimations of Clinical and Analytical Performance

Country	Brazil
Clinical Sensitivity (95% CI), N	74.4% (65.8, 81.4), 117
Sensitivity days ≤7, N	77.6% (68.3, 84.7), 98
Sensitivity Ct ≤ 33, N	82.5% (73.7, 88.8), 97
Sensitivity Ct ≤ 25, N	90.9% (80.5, 96.1), 55
Clinical Specificity (95% CI), N	98.95 (97.2, 99.6), 359
Invalid rate (% , n/N)	0% (0/476)
Analytical Sensitivity (pfu/ml)	5 x 10 ⁴ pfu/ml ~ 7.14 x 10 ⁴ TCID ₅₀ /ml (LSTM) ¹

Note: We found the verified LOD to be 20-35 times lower than the supplier reported LOD of 10^{3.54} – 10^{3.56} TCID₅₀/ml, which is the equivalent of 1.4-2.5x10³ pfu/ml, using a different viral strain.

5.3 Ease of Use

RapiGEN, Inc.	PENDING [XX] out of 100	[##] operators, [count(ies)]
---------------	-------------------------	------------------------------