

**FIND Evaluation of Guangzhou Wondfo Biotech Co., Ltd**  
**Wondfo 2019-nCoV Antigen Test (Lateral Flow Method)**  
**Public Report**  
*Version 1.0, 30 September 2021*

**Copyright and use of the report**

Copyright in this report is the property of FIND (or controlled by FIND). You are free to share, copy and redistribute the material in any medium or format provided that:

- (i) attribution: you must give appropriate credit to FIND and indicate if changes were made, you may do so in any reasonable manner, but not in any way that suggests that FIND endorses you or your use;
- (ii) non-commercial: you may not use the report for commercial purposes; and
- (iii) no derivatives: if you remix, transform, or build upon the materials or report, you may not distribute the modified materials or report unless with express authorization from FIND.

Presentation of data on our website does not impact any data ownership rights and FIND is not responsible for any use by any third party of these data. Data sources are provided.

**Evaluation process – private sector engagement**

FIND, the global alliance for diagnostics, seeks to ensure equitable access to reliable diagnosis around the world. It works closely with the private and public sectors and receives funding from donors and some of its industry partners. It has internal fire walls, policies and processes to protect it against any undue influence in its work or the publication of its findings.

More information on our policy and guidelines for working with private sector partners can be found here: <https://www.finddx.org/policies/>

For the COVID-19 response, FIND has commissioned independent evaluations of in vitro diagnostics following an Expression of Interest (EOI) process available on FIND’s webpage by which all test submissions were scored according to their regulatory status and time to market; the manufacturing and distribution capacity of the supplier; and the supplier-reported clinical and analytical performance.

**Document history**

Document Version	Date	Comment
1.0	30 September 2021	Initial version

## 1 Product info:

Manufacturer name	Guangzhou Wondfo Biotech Co., Ltd
Test name	Wondfo 2019-nCoV Antigen Test (Lateral Flow Method)
Product code(s)	W634P0013
Pack size(s)	20 tests/kit
Contents of kit	Test cassette, desiccant pouch, extraction tubes, drippers, sterile swabs, extraction buffer, IFU
Equipment and consumables required, but not provided	PPE, Timer, Biohazard container
Product storage (temperature range)	2-30°C
Shelf-life (months)	24 months
Manufacturing site (country)	China

## 2 Study details:

Study design:	Prospective diagnostic evaluation studies across multiple, independent sites to determine the accuracy of COVID-19 antigen RDTs, using consecutive enrolment. Interim analyses are performed at 25% and 50% enrolment, and the evaluation is stopped if tests do not meet 95% specificity. Presence of symptoms, date of symptom onset and hospitalization status is collected for all enrolled participants.
Index assays:	Novel lateral flow format tests that detect recombinant SARS-CoV-2 antigens.
Reference method:	Results of the index test are compared to the routine, diagnostic RT-PCR result, which is used for clinical management
Clinical performance:	<p>Sensitivity was calculated as the proportion of true positive results detected by Wondfo 2019-nCoV Antigen Test among all positives by the reference method and reported as a percentage.</p> <p>Specificity was calculated as the proportion of true negative specimens, identified as negative by Wondfo 2019-nCoV Antigen Test among all negatives by the reference method, and reported as a percentage.</p> <p>The 95% confidence intervals were calculated to assess the level of uncertainty introduced by sample size, using the Wilson's score method.</p>

### 3 Evaluation details

Country of collaborator	Brazil
Location of clinical site(s) (city, town)	1. Rio de Janeiro 2. Guapimirim State of Rio de Janeiro
Health care level of site(s)	1. Tertiary hospital 2. Community testing clinics
Study period (date to date)	17 August – 3 September 2021
Study cohort inclusion/exclusion	Children over 2 years old and adults in community meeting national suspect definition Provided informed consent or assent
Sample type, antigen test	Anterior nasal (Nasal) and Nasopharyngeal (NP)
Reference PCR method	Lab-developed assay based on the US CDC protocol, which targets two regions (N1 and N2) of the nucleocapsid (N) gene of SARS-CoV-2 ( <a href="https://www.fda.gov/media/134922/download">https://www.fda.gov/media/134922/download</a> )
Sample type, PCR test	Nasopharyngeal swab

### 4 Results

#### 4.1 Study cohort

<b>Country</b>	Brazil
Total N (valid PCR results)	237
Age [mean (min-max), N]	37.0 (3-107), 237
Gender [%F, (n/N)]	60.8% (144/237)
Positivity [%, (n/N)]	32% (76/237)
Symptoms present [%Yes, (n/N)]	99.6% (236/237)
Hospitalized (n, % Yes)	Not available
Days from symptom onset <sup>1</sup> [median (Q1-Q3); N]	4 (3-5); 236
Days < 0-3 (n, %)	98, 42%

Days 4-7 (n, %)	132, 56%
Days 8+ (n, %)	6, 3%
PCR Ct [median (Q1-Q3); N]	18.5 (14.4-23.8); 76
Ct > 33 (n, %)	1, 1%
Ct > 30 (n, %)	6, 8%
Ct > 25 (n, %)	15, 20%

<sup>1</sup>Note: data on symptom onset only available for individuals who tested PCR positive.

#### 4.2 Estimation of clinical performance

Country	Brazil	
	Nasal swab	Nasopharyngeal swab
Clinical Sensitivity (95% CI), N	88.2% (79, 93.6), 76	89.5% (80.6, 94.6), 76
Sensitivity days ≤7, N	89% (79.8, 94.3), 73	90.4% (81.5, 95.3), 73
Sensitivity Ct ≤ 33, N	88.0% (78.7, 93.6), 75	89.3% (80.3, 94.5), 75
Sensitivity Ct ≤ 25, N	96.7% (88.8, 99.1), 61	96.7% (88.8, 99.1), 61
Clinical Specificity (95% CI), N	98.1% (94.7, 99.4), 161	98.8% (95.6, 99.7), 161
Invalid rate (%), n/N	0% (0/237)	0% (0/237)
Positive percent agreement – nasal/NP (95% CI), N	98.6% (92.3, 99.7); 70	NA
Negative percent agreement – nasal/NP (95% CI), N	99.4% (96.7, 99.9); 167	NA