

FIND Evaluation of Rapigen
BIOCREDIT COVID-19 Ag
External Report
Version 1.0, [25 May 2022]

Copyright and use of the report

Copyright in this report is the property of FIND (or controlled by FIND). You are free to share, copy and redistribute the material in any medium or format provided that:

- (i) attribution: you must give appropriate credit to FIND and indicate if changes were made, you may do so in any reasonable manner, but not in any way that suggests that FIND endorses you or your use;
- (ii) non-commercial: you may not use the report for commercial purposes; and
- (iii) no derivatives: if you remix, transform, or build upon the materials or report, you may not distribute the modified materials or report unless with express authorization from FIND.

Presentation of data on our website does not impact any data ownership rights and FIND is not responsible for any use by any third party of these data. Data sources are provided.

Evaluation process – private sector engagement

FIND, the global alliance for diagnostics, seeks to ensure equitable access to reliable diagnosis around the world. It works closely with the private and public sectors and receives funding from donors and some of its industry partners. It has internal fire walls, policies and processes to protect it against any undue influence in its work or the publication of its findings.

More information on our policy and guidelines for working with private sector partners can be found here: <https://www.finddx.org/policies/>

For the COVID-19 response, FIND has commissioned independent evaluations of in vitro diagnostics following an Expression of Interest (EOI) process available on FIND’s website by which all test submissions were scored according to their regulatory status and time to market; the manufacturing and distribution capacity of the supplier; and the supplier-reported clinical and analytical performance.

Document history

Document version	Date	Comment
1.0	25 May 2022	First release

1 Product Info:

Manufacturer name	Rapigen Inc.
Test name	BIOCREDIT COVID-19 Ag
Product code(s)	Nasopharyngeal: G61RHA20 Nasal: G69RHA20
Pack size(s)	20 tests/kit
Contents of kit	Test device sealed in foil pouch with a desiccant, assay diluent tube, filter cap, sterilized swab, instructions for use
Equipment and consumables required, but not provided	PPE, timer
Product storage (temperature range)	2-30°C
Shelf-life (months)	24
Manufacturing site (country)	Republic of Korea

2 Study details:

Study design:	Prospective diagnostic evaluation studies across multiple, independent sites to determine the accuracy of COVID-19 antigen RDTs, using consecutive enrolment. Interim analyses are performed at 25% and 50% enrolment, and the evaluation is stopped if tests do not meet 95% specificity. Presence of symptoms, date of symptom onset and hospitalization status is collected for all enrolled participants.
Index assays:	Novel lateral flow format tests that detect recombinant SARS-CoV-2 antigens.
Reference method:	Results of the index test are compared to the routine, diagnostic RT-PCR result, which is used for clinical management
Limit of detection:	Analytical sensitivity, i.e. Limit of detection, was performed at the Liverpool School of Tropical Medicine in which standardized serial dilutions of cultured viral isolate were prepared. Proprietary swab provided in the kit was soaked in viral dilution series. Dilutions were tested in triplicate and the LOD was defined as the last dilution where all repeats were interpreted as positive.

Clinical performance:	<p>Sensitivity was calculated as the proportion of true positive results detected by Rapigen BIOCREDIT COVID-19 Ag among all positives by the reference method, and reported as a percentage</p> <p>Specificity was calculated as the proportion of true negative specimens, identified as negative by Rapigen BIOCREDIT COVID-19 Ag among all negatives by the reference method and reported as a percentage.</p> <p>The 95% confidence intervals were calculated to assess the level of uncertainty introduced by sample size, using the Wilson's score method.</p>
-----------------------	---

3 Evaluation details:

Country of collaborator	United Kingdom
Location of clinical site(s) (city, town)	Liverpool John Lennon Airport, United Kingdom
Health care level of site(s)	Drive through testing centre
Study period (date to date)	6 December 2021 – 18 March 2022
Study cohort inclusion/exclusion	<p>Inclusion:</p> <ul style="list-style-type: none"> • 18 years or older, who are undergoing testing for COVID-19 • Symptomatic for COVID-19 • Provided informed consent <p>Exclusion:</p> <ul style="list-style-type: none"> • Asymptomatic • Younger than 18 years
Sample type, antigen test	Nasopharyngeal and nasal swab
Reference PCR method	TaqPath™ COVID-19 CE-IVD RT-PCR Kit (ThermoFisher Scientific)
Sample type, PCR test	Nasopharyngeal swab

4 Results:

4.1 Study cohort

Country	United Kingdom
----------------	----------------

Total N (valid PCR results)	224
Age [mean (min-max), N]	42.8 (18-78), 224
Gender [%F, (n/N)]	58.5%, 131/224
Symptoms present [%Yes, (n/N)]	100%, 224/224
Hospitalized (n, % Yes)	Not applicable
Days from symptom onset [median (Q1-Q3); N]	2 (1-3), 224
Days < 0-3 (n, %)	179, 80%
Days 4-7 (n, %)	41, 18%
Days 8+ (n, %)	4, 2%
Positivity [%, (n/N)]	53%, 118/224
PCR Ct [median (Q1-Q3); N]	19.7 (17.0-24.7), 118
Ct > 33 (n, %)	5, 4%
Ct > 30 (n, %)	11, 9%
Ct > 25 (n, %)	28, 24%

4.2 Estimation of Clinical Performance

Country	United Kingdom	
	Nasopharyngeal	Nasal
Clinical Sensitivity (95% CI), N	82.2% (74.3, 88.1), 118	82.2% (74.3, 88.1), 118
Sensitivity days ≤7, N	82.1% (74.1, 88), 117	82.1% (74.1, 88), 117
Sensitivity Ct ≤ 33, N	85.8% (78.2, 91.1), 113	85.8% (78.2, 91.1), 113
Sensitivity Ct ≤ 25, N	92.2% (84.8, 96.2), 90	95.6% (89.1, 98.3), 90
Clinical Specificity (95% CI), N	98.1% (93.4, 99.5), 106	100% (96.5, 100), 106
Invalid rate (% , n/N)	0% (0/224)	0% (0/224)
Positive percent agreement – nasal/NP (95% CI), N	89.9% (82.4, 94.4), 99	NA
Negative percent agreement – nasal/NP	93.6% (87.9, 96.7), 125	NA

Important note: There was a 1h to 3h delay between Ag RDT sample collection and Ag RDT testing which may have degraded the overall performance results of the test.

4.2.1 Estimation of analytical performance

- Supplier-reported LOD

Limit of Detection	Specimen type	Name	Cat. No.	Source of strain
8.03 x 10² TCID₅₀/ml ~ 5.62 x 10 ² PFU/ml	Cell culture-derived virus	SARS-CoV-2 (S clade) (BetaCoV/Korea/KCDC03/2020)	43326 SARS-CoV-2	NCCP
5.35x10² TCID₅₀/ml ~ 3.74x10 ² PFU/ml	SARS-CoV-2 Inactivated virus (Heat inactivation)	Heat Inactivated 2019 Novel Coronavirus (USA-WA1/2020)	VR- 1986HK™	ATCC

- Verified LOD

Variant (lineage)	Lowest dilution detected	Verified LOD concentration	Viral Copy equivalent
UK wild type (B1)	5.0 x10³ pfu/ml ~ 7.05 x 10 ³ TCID ₅₀ /ml	5.0 x10³ pfu/ml	9.8x10⁶ genome copies/ml applied to test
Alpha (B.1.1.7)	5.0 x10⁴ pfu/ml ~ 7.05 x 10 ⁴ TCID ₅₀ /ml	5.0 x10³ pfu/ml	4.7 x10⁶ genome copies/ml applied to test
Gamma (P1)	1.0 x10³pfu/ml ~ 1.41 x 10 ³ TCID ₅₀ /ml	1.0 x10³ pfu/ml	5.60 x10⁵ genome copies/ml applied to test
Delta (B.1617.2)	1.0 x10² pfu/ml 1.41 x 10 ² TCID ₅₀ /ml	1.0 x10² pfu/ml	1.64 x10⁵ genome copies/ml applied to test
Omicron (BA.1)	2.5 x10² pfu/ml ~ 3.525 10 ² TCID ₅₀ /ml	2.5 x10² pfu/ml	8.80 x10⁴ genome copies/ml applied to test

Note: viral dilution was applied directly to the test cassette, not to the provided swab