

# Target product profile for a rapid diagnostic test for surveillance of cholera outbreaks

## Contents

List of abbreviations .....	1
Introduction.....	2
Recommendations for enhanced surveillance of cholera .....	2
Current state of play of cholera RDTs and molecular diagnostic tools.....	3
The role of target product profiles .....	3
Target product profile for a rapid diagnostic test for surveillance of cholera outbreaks.....	5
References .....	13

## List of abbreviations

GT FCC	Global Task Force on Cholera Control
IFU	instructions for use
LMICs	low- and middle-income countries
PAMI	priority area for multisectoral intervention
PCR	polymerase chain reaction
RDT	rapid diagnostic test
RUO	research use only
TPP	target product profile
WHO	World Health Organization

## Introduction

Cholera is a severe diarrhoeal disease caused by the bacterium *Vibrio cholerae*. It is a major threat to global public health, with an estimated 1.3 to 4 million cases and 21 000 to 143 000 deaths occurring annually (1). The average cholera case fatality ratio reported globally in 2021 was 1.9% (2.9% in Africa), substantially above the accepted targeted rate (< 1%) and the highest recorded in more than a decade (2). Approximately 10% of individuals with cholera will develop severe symptoms, including acute watery diarrhoea and vomiting; without treatment, death can occur within hours (3).

There is thus a considerable burden of cholera cases globally. In addition, the world has been facing an acute upsurge in what has been termed the seventh cholera pandemic; this is characterized by the number, size and concurrence of multiple outbreaks, the spread to areas free of cholera for decades and alarmingly high mortality rates (4). The simultaneous progression of multiple cholera outbreaks is compounded in countries that are experiencing complex humanitarian crises while relying on fragile health systems. This situation is further aggravated by climate change. The cholera pandemic thus presents a challenge to outbreak response, and there is the risk of the infection spreading further to other countries. Based on these factors, the increasing number of cholera outbreaks and their geographic expansion, in addition to the lack of vaccines and other relevant healthcare resources, the World Health Organization (WHO) assesses the current risk due to cholera globally to be very high (4).

Cholera disproportionately impacts the poorest and most vulnerable populations. Areas with poor sanitation, limited access to safe drinking water and deficient hygiene practices are at high risk for cholera transmission (3). In addition, limited access to healthcare facilities and inadequate treatment of cases are factors associated with a high level of cholera-related mortality.

The Global Task Force on Cholera Control (GTFCC) (5) and the WHO continue to work with partners, including Gavi (6), at the global, regional and country levels. This work involves supporting member states in their efforts to control cholera. It comprises a long-term, multisectoral approach that integrates various interventions, including improved sanitation infrastructure, surveillance, and the introduction of oral cholera vaccine as an outbreak response and for preventive vaccination campaigns.

## Recommendations for enhanced surveillance of cholera

Ending Cholera—A Global Roadmap to 2030 (5) (hereafter referred to as “the roadmap”) was launched by GTFCC in October 2017. The roadmap targets a 90% reduction in cholera deaths overall and elimination of cholera transmission in 20 countries by 2030. In 2018, the roadmap was recognized by the 21st World Health Assembly and adopted by the WHO Regional Committee for Africa.

The strategy calls on countries and partners to focus on cholera “hotspots” or priority areas for multisectoral intervention (PAMIs). These are relatively small areas that are most heavily affected by cholera; they experience cases on an ongoing or seasonal basis and play a major role in the spread of cholera to other areas. The roadmap is based on three strategic axes:

- (1) Early detection and quick response to rapidly contain outbreaks of cholera.
- (2) Implementation of a multisectoral approach targeted at high-risk areas. This includes the delivery of oral cholera vaccines; basic water, sanitation and hygiene (WASH) services; epidemiology and laboratory services; case management; and community engagement.
- (3) Effective coordination of technical support, resource mobilization, and partnership at local and global levels.

All three strategic axes of the roadmap are heavily dependent on the availability of reliable, accurate and high-quality epidemiological and laboratory surveillance data. Such data are necessary to detect, confirm and rapidly respond to cholera outbreaks; identify high-risk areas; monitor progress; assess (and adapt) prevention and control strategies; and substantiate the absence of cholera. The rapid identification of

cholera PAMIs, based on existing surveillance data, is essential to tailor the correct combination of interventions that are best adapted to each context and population.

The objective of cholera surveillance and control strategies has long been limited to responding to and mitigating major epidemics. As a result, cholera surveillance has mainly relied on the reporting of clinically suspected cases and deaths, with no or limited laboratory confirmation. Therefore, the currently available cholera surveillance data are not sufficiently reliable to adequately support the objectives of the roadmap or to identify PAMIs across and within countries. Not all actual cases of cholera are reported; conversely, not all clinically suspected cholera cases are true cases, as the clinical signs of cholera are also commonly seen with other frequently occurring diarrhoeal diseases.

In February 2023, GTFCC published updated recommendations for strengthening public health surveillance for cholera, with the aim of having better informed, more timely and appropriately targeted multisectoral interventions (7). The guidance states that laboratory testing should rely on testing strategies adapted to the prevailing cholera situation at the surveillance unit level (i.e. the presence or absence of a confirmed cholera outbreak) and to available resources; expanded use of rapid diagnostic tests (RDTs) to support the early detection of suspected outbreaks and incidence monitoring; and increased capacities for laboratory confirmation of cholera cases. Testing of suspected cholera cases should be routinely undertaken in accordance with systematic testing schemes, and again the use of RDTs should be expanded.

## Current state of play of cholera RDTs

RDTs are primarily intended to be used at the primary healthcare level for surveillance purposes. They may be used as a tool for triaging samples before possible further testing in laboratories; to speed up outbreak detection in surveillance units where a cholera outbreak has yet to be confirmed; and to help monitor incidence trends of actual cholera cases in surveillance units with a confirmed cholera outbreak, through systematic testing of a representative proportion of suspected cases (7).

Although there are cholera RDTs on the market, most are designated for research use only (RUO), and a few are self-declared CE-marked in vitro diagnostic (IVD) devices; none have been prequalified by WHO. While the performances of most of these cholera RDTs have been evaluated and the results published in the scientific literature, these studies have used a variety of non-standardized approaches, meaning they are often difficult to compare. In the absence of a prequalified product, WHO and UNICEF have procured cholera RDTs (8) for use in outbreaks, based on their own selection protocols and reviews of these products. To improve the quality and availability of laboratory data for cholera and ultimately reach the targets set out in the roadmap, there is an urgent need for easy-for-use, fit-for-purpose, well-performing and validated diagnostic tests.

## The role of target product profiles

Target product profiles (TPPs) are strategic planning tools for guiding the development of new diagnostic tests and other healthcare products. The primary audiences for TPPs are manufacturers, suppliers and researchers developing new assays. A TPP outlines the key performance and operational characteristics that a product should possess to meet the needs of its intended users, target population and public health programmes, in their intended settings of use. For each characteristic, a TPP states a preferred criterion that is to be achieved by product developers if feasible and a minimal criterion if the preferred criterion is not feasible.

Since 2017, GTFCC has focused on encouraging the availability and use of RDTs, developing the TPP for cholera RDTs first published in 2017 (9) and related WHO prequalification standards (10). However, since 2017, GTFCC has agreed to modify some of the previously agreed upon minimal characteristics that were deemed too stringent, thus there is a need to update the TPP accordingly. The TPP for a cholera RDT described in this document is informed by the previously developed TPP, but here we incorporate advances in the field and the most recently updated GTFCC guidelines (7). Importantly, this document has two intended goals to support cholera surveillance, relating to the minimal and the preferred criteria.

The goal of the minimal test criterion addresses the use of RDTs for detecting outbreak alerts and monitoring ongoing outbreaks, in accordance with current GTFCC guidelines, which require identification of at least *V. cholerae* O1 with performance requirements that are likely attainable based on what is known for currently available best-in-class commercial tests (7). It is assumed that the detection of *V. cholerae* O139 is less important due to the absence of O139 outbreaks over the past decade; in fact, false-positive O139 test results have become a challenge for both countries and surveillance units. Therefore, detection of O139 is only included in the preferred criterion for assay targets.

A test that meets the preferred test goal, although potentially challenging to achieve, would have the potential to supplant the current requirement for case confirmation by polymerase chain reaction (PCR) or culture through addressing the preferred criteria of additional assay targets (O139 and cholera toxin) and improved sensitivity and specificity similar to that of PCR or culture. Having an ambitious preferred test goal serves to drive innovations that would substantially improve cholera response efforts, by enabling faster and less costly cholera outbreak confirmation through the decentralization of case confirmation and removing the costs of sample transport and complex laboratory testing to confirm an outbreak. If an RDT can meet these preferred criteria at the prices described in this TPP, it would be more attractive to those programmes and procurement agencies supporting cholera responses compared with a test that only satisfies the minimal criteria. The window is currently open for manufacturers with cholera RDTs that meet the TPP to submit expressions of interest to WHO for assessment of their products for prequalification.

## Target product profile for a rapid diagnostic test for surveillance of cholera outbreaks

Please note that the numbering of notes restarts after each row.

Characteristic	Minimal	Preferred	Notes
Goal of the test	A cholera surveillance test <sup>1</sup> to detect outbreak alerts and to monitor ongoing outbreaks. <sup>2</sup>	Same as minimal, plus the ability to confirm cholera outbreaks. <sup>3</sup>	<p><sup>1</sup> According to WHO and the International Medical Device Regulators Forum (IMDRF) categories of test purposes (11). The tests described under both the minimal and preferred criteria are for surveillance purposes; they are not meant to inform patient management decisions, only to identify cholera cases.</p> <p><sup>2</sup> According to current guidelines, culture and/or PCR is required to confirm an individual case of cholera or confirm a cholera outbreak. Culture and/or PCR will likely need to be performed at a reference laboratory with the necessary laboratory capacity.</p> <p><sup>3</sup> Test to be used on a predefined number of suspected cholera cases to enable case detection and declaration of an outbreak without the need for further laboratory confirmation by PCR/culture.</p>
Target population	All patients ≥ 2 years old who meet the clinical definition of a suspected case of cholera (7).		

Characteristic	Minimal	Preferred	Notes
Target use setting	Primary healthcare settings, including health posts (level 1 and above). <sup>1</sup>		<sup>1</sup> For definitions of healthcare levels, see Table 1.
Assay targets	Biomarkers for detection of <i>Vibrio cholerae</i> O1 (e.g. lipopolysaccharide (LPS)-O antigens) to enable surveillance for outbreak alert and monitoring.	Same as minimal, plus detection and differentiation of <i>V. cholerae</i> O139 and detection of cholera toxin to identify strains known to cause epidemics. <sup>1</sup>	<sup>1</sup> Assay targets may be in separate tests or, preferably, combined in a single test. To satisfy the preferred goal of the test, i.e. outbreak confirmation, detection of cholera toxin is required.
Test format	A single-use, disposable test kit that requires no instruments or laboratory equipment to perform the test procedure, including specimen preparation.  The test kit includes all materials required for the test procedure to test one individual, including devices, reagents and other consumables, in a packaged, self-contained kit. Additional consumables may be needed for specimen collection. (A reader as an optional tool for interpreting results is acceptable.)	Same as minimal, plus consumables required to support specimen collection included in the test kit.	
Result display and interpretation	Visual interpretation of qualitative (yes/no) test results by the naked eye, with minimal instructions for interpretation by the user required. (In addition, results could be obtained via an optional reader. <sup>1</sup> )		<sup>1</sup> Characteristics of a reader are defined in a WHO TPP for readers of RDTs (12).

DRAFT for 1<sup>st</sup> public Delphi survey

Characteristic	Minimal	Preferred	Notes
User requirements			
Target users	Community health workers with minimal training and any healthcare provider with similar or superior training.	Same as minimal, plus trained health workers from outbreak investigation teams present at the community level and community health workers at primary or secondary care levels.	
User training	A user can conduct the test correctly after half a day of training.	A user can conduct the test correctly after a brief review of the instructions for use (IFU).	
Performance characteristics			
Specimen type	Unprocessed stool specimen or faecal swab (the test must support both so that the user can choose which to use).	Same as minimal, or a specimen that is easier to collect, e.g. capillary blood, if new biomarkers have been identified and validated.	
Clinical sensitivity <sup>1</sup>	≥ 90% for each assay target <sup>2</sup>	≥ 95% for each assay target <sup>3</sup>	<p><sup>1</sup> For performance studies, see (10)</p> <p><sup>2</sup> As recommended by the GTFCC laboratory working group (13)</p> <p><sup>3</sup> To address the preferred goal of the test for outbreak confirmation, this preferred sensitivity level is required, which is equivalent to the minimal performance requirement of molecular tests for cholera outbreak confirmation.</p>

Characteristic	Minimal	Preferred	Notes
Clinical specificity <sup>1</sup>	≥ 85% for each assay target <sup>2</sup>	≥ 95% for each assay target <sup>3</sup>	<p><sup>1</sup> For performance studies, see (10)</p> <p><sup>2</sup> As recommended by the GTFCC laboratory working group (13)</p> <p><sup>3</sup> To address the preferred goal of the test for outbreak confirmation, this preferred specificity level is required.</p>
Limit of detection	To be determined <sup>1</sup>		<sup>1</sup> Further research is needed to define the clinical reference ranges of bacterial load among infected individuals and to certify reference materials for the assay targets.
Analytical specificity and assay interference	No cross-reactivity with <i>V. cholerae</i> non-O1/non-O139 or with biomarkers of common human diseases, especially those presenting with similar signs and symptoms, e.g. watery diarrhoea, or with common endogenous or exogenous interferents. <sup>1</sup>		<sup>1</sup> Pathogens and interferents should be tested at clinically relevant concentrations, included in a risk evaluation and listed in the IFU (see (10)).
Test failure (invalid) rate	≤ 5%	≤ 1%	
Test procedure characteristics			
Specimen volume	< 1 mL for stool specimens; < 0.1 mL for faecal swabs	Same as minimal, or small volumes consistent with the specimen type collected, e.g. 25–50 µL for capillary blood.	



DRAFT for 1<sup>st</sup> public Delphi survey

Characteristic	Minimal	Preferred	Notes
Specimen preparation	<p>None required other than swab elution or dilution steps where no precision pipetting is required.</p> <p>The test may support an optional step of enrichment in alkaline peptone water (APW).<sup>1</sup></p>	None required other than swab elution or dilution steps where no precision pipetting is required.	<sup>1</sup> The test must satisfy the minimal clinical specificity characteristic without APW enrichment. If the test supports the option of enrichment for higher specificity, the IFU must provide instructions and summarized results of performance studies with and without enrichment.
Ease of use	Easy-to-perform test procedure and result interpretation by the intended user, with minimal steps; no precision pipetting and no timed steps (except for reading the test result). Reagent reconstitution is acceptable if simple to do and if all liquids, including water, are provided in the test kit.	Same as minimal, except no reagent reconstitution	See (10)
Language support	For each country of deployment, the packaging and IFU are provided in one commonly used language, such as the official language or the de facto national language, and any language mandated by local regulatory or trade compliance requirements.	Same as minimal, plus additional languages to enable use by other residents of that country.	Relevant countries for cholera RDTs are likely to use Arabic, English, French, Portuguese or Swahili.
Time to result <sup>1</sup>	≤ 30 min	≤ 15 min	<sup>1</sup> Including time from the start of specimen preparation to test result, not including the duration of specimen collection.
Stability of valid result	≥ 30 min (after which results may be false or invalid).	≥ 1 h (after which results give invalid rather than false results).	

DRAFT for 1<sup>st</sup> public Delphi survey

Characteristic	Minimal	Preferred	Notes
Internal controls	Control of reagent addition and flow	Same as minimal, plus control of specimen addition and flow, of correct operation of the device, and of correct functionality of all reagents.	
External controls	Positive and negative controls are specified in the IFU and available for purchase separately.	Positive and negative controls are included in the price of the test and are provided with the test kits. <sup>1</sup>	<sup>1</sup> For logistics and stability, controls may be packaged separately from the test kits.
Operational characteristics			
Test kit stability and operational conditions (14)	18 months, stable at 2–30 °C and 70% humidity; any associated equipment must meet or exceed these requirements. <sup>1</sup>	24 months, stable at 2–40 °C and 90% humidity; any associated equipment must meet or exceed these requirements. <sup>1</sup>	<sup>1</sup> Based on specifications in the WHO TPP for a point-of-care test for prior infection with SARS-CoV-2 (15).
Shipping and transport conditions (14)	Temperature stress (48 h with fluctuations up to 40 °C and down to 2 °C); no cold chain required for storage or transport.	Same as minimal, except temperature stress (72 h with fluctuations up to 50 °C and down to 2 °C), and with indicator of temperature or humidity exceedances that would result in invalid or low-performance results.	
Open test kit stability <sup>1</sup>	≥ 30 min	≥ 1 h	<sup>1</sup> Including time from opening the test kit to finishing the addition of specimen and reagents.
Biosafety	None, apart from waste management and the use of non-sterile gloves.		

Characteristic	Minimal	Preferred	Notes
Waste disposal	Standard biohazard waste disposal or incineration of consumables; no high-temperature incineration required.	Test cassettes and consumables designed to minimize environmental footprint during standard biohazard waste disposal.	
Pricing and market access			
Target list price per test <sup>1</sup>	A test meeting the minimal goal of the test (for surveillance for outbreak alert and monitoring): < US\$ 3.	<p>A test meeting the minimal goal of the test (for surveillance for outbreak alert and monitoring): &lt; US\$ 1.</p> <p>A test meeting the preferred goal of the test (capable of outbreak confirmation): &lt; US\$ 3.</p>	<p><sup>1</sup> Pricing from manufacturers should be as low as sustainably possible while maintaining quality, based on evidence of the true cost of goods sold accounting for material, manufacturing process, operational logistics and commercialization efforts. Pricing should also include and clearly define all facets of end-to-end implementation (e.g. support, maintenance). Pricing must account for production at scale with defined volume thresholds. Ultimately pricing should intersect sustainable long-term viability for the manufacturer with affordability to support widespread access to testing in LMICs and should be transparently published.</p> <p>This price excludes any costs for a reader.</p>
Quality management system	Compliant with ISO 13485	Certified to ISO 13485 or equivalent	
Lead time for lot production	≤ 3 months	≤ 1 month	

Table 1: Definitions of use settings in LMICs (adapted from (16))

	Self-Testing	Level 0 (L0) - Community	Level 1 (L1) - Primary Care	Level 2 (L2) - District Hospital Lab	Level 3 (L3) - Regional/Provincial Lab	Level 4 (L4) – Reference/National Lab
<b>Use setting</b>	<ul style="list-style-type: none"> <li>• Home testing</li> </ul>	<ul style="list-style-type: none"> <li>• Community outreach</li> <li>• Home testing</li> </ul>	<ul style="list-style-type: none"> <li>• Primary care facility</li> </ul>	<ul style="list-style-type: none"> <li>• Near-patient laboratory</li> <li>• Referral hospital laboratory</li> <li>• Emergency Department testing</li> </ul>	<ul style="list-style-type: none"> <li>• Near-patient laboratory</li> <li>• Referral hospital laboratory</li> <li>• Emergency Department testing</li> </ul>	<ul style="list-style-type: none"> <li>• Reference laboratory</li> </ul>
<b>Lab infrastructure</b>	<ul style="list-style-type: none"> <li>• No mains power</li> <li>• No water</li> <li>• No lab equipment</li> <li>• No environmental control (e.g., temp, dust, humidity)</li> </ul>	<ul style="list-style-type: none"> <li>• No mains power</li> <li>• No water</li> <li>• No lab equipment</li> <li>• No environmental control (e.g., temp, dust, humidity)</li> </ul>	<ul style="list-style-type: none"> <li>• No mains power (unreliable)</li> <li>• Minimal lab equipment (may not support cold chain)</li> <li>• BSL-1 containment</li> <li>• No environmental control (e.g., temp, dust, humidity)</li> </ul>	<ul style="list-style-type: none"> <li>• Mains power (may be intermittent)</li> <li>• Basic lab equipment (biosafety cabinet, centrifuge, calibrated pipets, fridge)</li> <li>• -20 freezers (some)</li> <li>• BSL-2/1 containment (some)</li> <li>• Environmental control (e.g., temp, dust, humidity) (some)</li> </ul>	<ul style="list-style-type: none"> <li>• Mains power (may be intermittent)</li> <li>• Basic lab equipment (biosafety cabinet, centrifuge, calibrated pipets, fridge)</li> <li>• -20 freezers</li> <li>• BSL-2/1 containment</li> <li>• Environmental control (e.g., temp, dust, humidity)</li> </ul>	<ul style="list-style-type: none"> <li>• Mains power (reliable)</li> <li>• High infrastructure facility</li> <li>• -20 freezers</li> <li>• -80 freezers (some)</li> <li>• BSL-2/3 containment</li> <li>• Environmental control (e.g., temp, dust, humidity)</li> </ul>
<b>Operator skill</b>	<ul style="list-style-type: none"> <li>• Self-testing</li> <li>➢ Simple reagent/sample transfer</li> </ul>	<ul style="list-style-type: none"> <li>• Nurse/pharmacist</li> <li>• Community health workers</li> <li>➢ Simple reagent/sample transfer</li> </ul>	<ul style="list-style-type: none"> <li>• Nurse</li> <li>• Trained laboratory worker</li> <li>➢ Minimal sample processing (≤ 3 steps)</li> </ul>	<ul style="list-style-type: none"> <li>• Laboratory technician (1-2 year certif)</li> <li>➢ Sample processing with calibrated volumes (≤ 3 steps)</li> </ul>	<ul style="list-style-type: none"> <li>• Laboratory technician (1-2 year certif)</li> <li>➢ Sample processing with calibrated volumes (≤ 3 steps)</li> </ul>	<ul style="list-style-type: none"> <li>• Science research specialists</li> <li>• Laboratory technician (1-2 year certif)</li> </ul>
<b>Specimen capacity</b>	<ul style="list-style-type: none"> <li>• Can process minimally invasive samples: fingerstick blood, nasal swabs, saliva, urine</li> </ul>	<ul style="list-style-type: none"> <li>• Can process minimally invasive samples: fingerstick blood, nasal swabs, saliva, urine</li> </ul>	<ul style="list-style-type: none"> <li>• Can process upper respiratory specimens; clinic may not have capacity for lower respiratory, venipuncture, plasma</li> </ul>	<ul style="list-style-type: none"> <li>• Can process most BSL-2 specimens; depends on clinic sample capacity</li> </ul>	<ul style="list-style-type: none"> <li>• Can process most BSL-2 specimens; depends on clinic sample capacity</li> </ul>	<ul style="list-style-type: none"> <li>• Can process most BSL2/3 specimens</li> </ul>
<b>Test capacity</b>	<ul style="list-style-type: none"> <li>• True-POC MDx (some)</li> <li>• RDT</li> </ul>	<ul style="list-style-type: none"> <li>• True-POC MDx (some)</li> <li>• RDT</li> </ul>	<ul style="list-style-type: none"> <li>• True-POC MDx</li> <li>• Basic microscopy</li> <li>• RDT</li> </ul>	<ul style="list-style-type: none"> <li>• Near-POC MDx</li> <li>• ELISA with simple reader</li> <li>• Microscopy</li> <li>• RDT</li> <li>• Clinical chemistry (some)</li> </ul>	<ul style="list-style-type: none"> <li>• Blood culture and microbiology capacity (some)</li> <li>• Near-POC MDx</li> <li>• ELISA with simple reader</li> <li>• Microscopy</li> <li>• RDT</li> <li>• Clinical chemistry</li> </ul>	<ul style="list-style-type: none"> <li>• Blood culture and microbiology capacity</li> <li>• Lab MDx / PCR / LDT</li> <li>• ELISA/EIA/CLIA/PRNT</li> <li>• Fluorescence microscopy</li> <li>• Clinical chemistry</li> <li>• Sequencing (some)</li> <li>• Mass spectrometry (some)</li> </ul>

BSL, biosafety level; CLIA, chemiluminescent assay; EIA, enzyme immunoassay; ELISA, enzyme-linked immunosorbent assay; LDT, laboratory developed test; MDx, molecular diagnostic test; POC, point of care; PRNT, plaque reduction neutralization test; RDT, rapid diagnostic test.

## References

1. Ali M, Nelson AR, Lopez AL, Sack DA. Updated global burden of cholera in endemic countries. PLoS Negl Trop Dis. 2015;9:e0003832. doi: 10.1371/journal.pntd.0003832.
2. World Health Organization. Cholera – Global situation. 2022 (<https://www.who.int/emergencies/disease-outbreak-news/item/2022-DON426>, accessed 29 November 2023).
3. Centers for Disease Control and Prevention. Cholera - Vibrio cholerae infection. 2023 (<https://www.cdc.gov/cholera/general/index.html#one>, accessed 24 October 2023).
4. World Health Organization. Multi-country outbreak of cholera, External Situation Report #1. 2023 ([https://cdn.who.int/media/docs/default-source/documents/emergencies/20230320\\_multi-country\\_outbreak-of-cholera\\_sitrep-1.pdf?sfvrsn=eaeed11a\\_1&download=true](https://cdn.who.int/media/docs/default-source/documents/emergencies/20230320_multi-country_outbreak-of-cholera_sitrep-1.pdf?sfvrsn=eaeed11a_1&download=true), accessed 29 November 2023).
5. Global Task Force on Cholera Control. Ending Cholera—A Global Roadmap to 2030. 2017 (<https://www.gtfcc.org/wp-content/uploads/2019/10/gtfcc-ending-cholera-a-global-roadmap-to-2030.pdf>, accessed 29 November 2023).
6. Gavi. Oral cholera vaccine support. nd (<https://www.gavi.org/types-support/vaccine-support/oral-cholera>, accessed 24 October 2023).
7. Global Taskforce on Cholera Control. Public health surveillance for cholera - interim guidance. 2023 (<https://www.gtfcc.org/wp-content/uploads/2023/02/gtfcc-public-health-surveillance-for-cholera-interim-guidance.pdf>, accessed 29 November 2023).
8. UNICEF. Cholera kits. nd (<https://supply.unicef.org/all-materials/in-vitro-diagnostic-tests/cholera-kits.html>, accessed 30 October 2023).
9. Global Task Force on Cholera Control. Target product profile (TPP) for the development of improved cholera rapid diagnostic tests. 2017 (<https://www.gtfcc.org/wp-content/uploads/2019/10/gtfcc-tpp-for-the-development-of-improved-cholera-rapid-diagnostic-test.pdf>, accessed 29 November 2023).
10. World Health Organization. Technical Specifications Series for submission to WHO Prequalification – Diagnostic Assessment: Rapid diagnostic tests (RDTs) used for surveillance and detection of an outbreak of cholera. 2018 (<https://iris.who.int/bitstream/handle/10665/260469/9789241513715-eng.pdf?sequence=1>, accessed 29 November 2023).
11. World Health Organization. The selection and use of essential in vitro diagnostics: report of the third meeting of the WHO Strategic Advisory Group of Experts on In Vitro Diagnostics, 2020 (including the third WHO model list of essential in vitro diagnostics). 2021 (<https://iris.who.int/handle/10665/339064>, accessed 29 November 2023).
12. World Health Organization. Target product profile for readers of rapid diagnostic tests. 2023 (<https://iris.who.int/bitstream/handle/10665/365980/9789240067172-eng.pdf?sequence=1>, accessed 29 November 2023).
13. World Health Organization. Interim technical note - the use of cholera rapid diagnostic tests. 2016 (<https://www.gtfcc.org/wp-content/uploads/2019/10/gtfcc-interim-use-of-cholera-rapid-diagnostic-tests.pdf>, accessed 29 November 2023).

14. World Health Organization. WHO Expert Committee on Biological Standardization: sixty-eighth report. 2017 (<https://iris.who.int/bitstream/handle/10665/272807/9789241210201-eng.pdf?sequence=1#page=52&zoom=100,0,0>, accessed 29 November 2023).
15. World Health Organization. TPP: For a point of care test for prior infection with SARS-CoV-2. 2020 (<https://www.who.int/tools/target-product-profile-database/item/tpp--for-a-point-of-care-test-for-prior-infection-with-sars-cov-2>, accessed 24 October 2023).
16. Ghani AC, Burgess DH, Reynolds A, Rousseau C. Expanding the role of diagnostic and prognostic tools for infectious diseases in resource-poor settings. *Nature*. 2015;528:S50-2. doi: 10.1038/nature16038.